

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111424\
 Data File : VU061685.D
 Acq On : 14 Nov 2024 21:02
 Operator : MD/SY
 Sample : VU1114WBL02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK085

Quant Time: Nov 15 03:21:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 15 03:17:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	185332	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	172893	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	73684	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	48006	4.009	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.200%
7) Chloroethane-d5	1.911	69	44203	4.324	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.400%
11) 1,1-Dichloroethene-d2	2.560	65	20333	3.785	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.600%
20) 2-Butanone-d5	4.631	46	128294	50.477	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.960%
24) Chloroform-d	5.052	84	104493	4.390	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
26) 1,2-Dichloroethane-d4	5.692	65	55436	4.853	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.000%
32) Benzene-d6	5.718	84	209311	4.376	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
36) 1,2-Dichloropropane-d6	6.679	67	65809	4.665	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.200%
41) Toluene-d8	7.891	98	171808	3.854	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.000%
43) trans-1,3-Dichloroprop...	8.174	79	20821	4.007	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%
46) 2-Hexanone-d5	8.627	63	105711	51.482	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.960%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	52662	4.735	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	56416	4.299	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111424\
Data File : VU061685.D
Acq On : 14 Nov 2024 21:02
Operator : MD/SY
Sample : VU1114WBL02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK085

Quant Time: Nov 15 03:21:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 15 03:17:58 2024
Response via : Initial Calibration

