

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111424\
 Data File : VU061695.D
 Acq On : 15 Nov 2024 01:31
 Operator : MD/SY
 Sample : P4747-02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 15 03:24:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 15 03:17:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	178454	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	165379	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	70445	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	51631	4.478	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.600%
7) Chloroethane-d5	1.911	69	52077	5.291	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.800%
11) 1,1-Dichloroethene-d2	2.560	65	23550	4.552	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.000%
20) 2-Butanone-d5	4.627	46	135221	55.253	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.500%
24) Chloroform-d	5.052	84	112169	4.894	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.800%
26) 1,2-Dichloroethane-d4	5.692	65	58205	5.292	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.800%
32) Benzene-d6	5.717	84	232220	5.076	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
36) 1,2-Dichloropropane-d6	6.679	67	71518	5.300	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.000%
41) Toluene-d8	7.891	98	200285	4.697	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
43) trans-1,3-Dichloroprop...	8.174	79	22117	4.450	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.000%
46) 2-Hexanone-d5	8.630	63	116235	59.179	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.360%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	56203	5.283	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	69422	5.533	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111424\
Data File : VU061695.D
Acq On : 15 Nov 2024 01:31
Operator : MD/SY
Sample : P4747-02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 15 03:24:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 15 03:17:58 2024
Response via : Initial Calibration

