

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111724\
 Data File : VU061782.D
 Acq On : 17 Nov 2024 14:18
 Operator : MD/SY
 Sample : P4800-19
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCPB8

Quant Time: Nov 18 00:05:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 18 00:04:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	234591	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	220804	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.805	152	99776	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	44567	2.940	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	58.800%
7) Chloroethane-d5	1.908	69	51744	3.999	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.000%
11) 1,1-Dichloroethene-d2	2.561	65	22490	3.307	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.200%
20) 2-Butanone-d5	4.625	46	133449	41.481	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.960%
24) Chloroform-d	5.052	84	117477	3.899	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.000%
26) 1,2-Dichloroethane-d4	5.695	65	58808	4.067	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.400%
32) Benzene-d6	5.718	84	239001	3.913	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.200%
36) 1,2-Dichloropropane-d6	6.683	67	73235	4.065	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.200%
41) Toluene-d8	7.891	98	207653	3.647	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.000%
43) trans-1,3-Dichloroprop...	8.174	79	22021	3.318	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.400%
46) 2-Hexanone-d5	8.628	63	115699	44.120	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.240%
56) 1,1,2,2-Tetrachloroeth...	10.747	84	58473	4.117	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	72560	4.083	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.600%
Target Compounds					Qvalue	
13) Acetone	2.638	43	123937	61.147	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111724\
Data File : VU061782.D
Acq On : 17 Nov 2024 14:18
Operator : MD/SY
Sample : P4800-19
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCPB8

Quant Time: Nov 18 00:05:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 18 00:04:27 2024
Response via : Initial Calibration

