

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111724\
 Data File : VU061792.D
 Acq On : 17 Nov 2024 18:24
 Operator : MD/SY
 Sample : P4829-15
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E29S2

Quant Time: Nov 18 00:07:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 18 00:04:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	169556	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	162387	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	67548	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	39652	3.620	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.400%
7) Chloroethane-d5	1.907	69	40035	4.281	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.600%
11) 1,1-Dichloroethene-d2	2.560	65	18828	3.830	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.600%
20) 2-Butanone-d5	4.621	46	114983	49.449	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.900%
24) Chloroform-d	5.052	84	97848	4.493	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
26) 1,2-Dichloroethane-d4	5.692	65	50896	4.870	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.400%
32) Benzene-d6	5.718	84	199577	4.443	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	6.682	67	64353	4.857	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.200%
41) Toluene-d8	7.891	98	172564	4.121	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.400%
43) trans-1,3-Dichloroprop...	8.174	79	18423	3.775	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.400%
46) 2-Hexanone-d5	8.627	63	98982	51.324	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.640%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	50152	4.801	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	60498	5.029	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.600%
Target Compounds					Qvalue	
16) Methylene chloride	3.039	84	6455	0.547	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111724\
Data File : VU061792.D
Acq On : 17 Nov 2024 18:24
Operator : MD/SY
Sample : P4829-15
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E29S2

Quant Time: Nov 18 00:07:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 18 00:04:27 2024
Response via : Initial Calibration

