

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111724\
 Data File : VU061783.D
 Acq On : 17 Nov 2024 14:42
 Operator : MD/SY
 Sample : VIBLK173
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK173

Quant Time: Nov 18 00:05:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 18 00:04:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	160705	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	154767	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	68073	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	57043	5.494	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.800%
7) Chloroethane-d5	1.911	69	51297	5.787	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.800%
11) 1,1-Dichloroethene-d2	2.560	65	25291	5.429	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	108.600%
20) 2-Butanone-d5	4.628	46	132639	60.184	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.360%
24) Chloroform-d	5.052	84	115553	5.599	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.000%
26) 1,2-Dichloroethane-d4	5.692	65	60702	6.129	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	122.600%
32) Benzene-d6	5.718	84	241882	5.649	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.000%
36) 1,2-Dichloropropane-d6	6.679	67	74150	5.872	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.400%
41) Toluene-d8	7.891	98	214649	5.379	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%
43) trans-1,3-Dichloroprop...	8.174	79	23859	5.129	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.600%
46) 2-Hexanone-d5	8.627	63	109671	59.666	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	119.340%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	57857	5.812	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	116.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	72007	5.939	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111724\
Data File : VU061783.D
Acq On : 17 Nov 2024 14:42
Operator : MD/SY
Sample : VIBLK173
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK173

Quant Time: Nov 18 00:05:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 18 00:04:27 2024
Response via : Initial Calibration

