

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111724\
 Data File : VU061785.D
 Acq On : 17 Nov 2024 15:32
 Operator : MD/SY
 Sample : P4829-21
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E29W2

Quant Time: Nov 18 00:06:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 18 00:04:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	198718	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	185059	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	80339	5.000	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.595	65	41481	3.231	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.600%
7) Chloroethane-d5	1.907	69	43580	3.976	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.600%
11) 1,1-Dichloroethene-d2	2.560	65	21048	3.654	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.000%
20) 2-Butanone-d5	4.618	46	124122	45.546	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.100%
24) Chloroform-d	5.052	84	107316	4.205	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
26) 1,2-Dichloroethane-d4	5.692	65	55956	4.569	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.400%
32) Benzene-d6	5.717	84	218720	4.272	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.400%
36) 1,2-Dichloropropane-d6	6.679	67	69763	4.620	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.891	98	190612	3.994	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
43) trans-1,3-Dichloroprop...	8.174	79	20739	3.729	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.600%
46) 2-Hexanone-d5	8.627	63	110054	50.074	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.140%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	53562	4.500	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	67289	4.703	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111724\
Data File : VU061785.D
Acq On : 17 Nov 2024 15:32
Operator : MD/SY
Sample : P4829-21
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E29W2

Quant Time: Nov 18 00:06:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 18 00:04:27 2024
Response via : Initial Calibration

