

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112022\
 Data File : VU051981.D
 Acq On : 20 Nov 2022 18:14
 Operator : JC/MD
 Sample : N5710-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0KY4

Quant Time: Nov 21 01:03:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM111522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 18 00:53:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	429030	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	427449	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	152128	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	79527	37.758	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	75.520%		
7) Chloroethane-d5	1.906	69	92731	43.942	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	87.880%		
11) 1,1-Dichloroethene-d2	2.565	63	102634	27.173	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	54.340%#		
21) 2-Butanone-d5	4.616	46	232740	97.176	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	97.180%		
24) Chloroform-d	5.060	84	240712	45.590	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.180%		
26) 1,2-Dichloroethane-d4	5.700	65	152362	47.141	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.280%		
32) Benzene-d6	5.726	84	412685	46.304	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.600%		
36) 1,2-Dichloropropane-d6	6.687	67	160984	48.526	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	97.060%		
41) Toluene-d8	7.896	98	302145	41.302	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	82.600%		
43) trans-1,3-Dichloroprop...	8.179	79	53724	40.657	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	81.320%		
47) 2-Hexanone-d5	8.629	63	147653	87.574	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	87.570%		
56) 1,1,2,2-Tetrachloroeth...	10.754	84	302510	46.852	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	93.700%		
66) 1,2-Dichlorobenzene-d4	12.192	152	147936	55.258	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	110.520%		
Target Compounds						
					Qvalue	
13) Acetone	2.620	43	93720	36.575	ug/L	99
15) Methyl Acetate	2.951	43	4193	1.200	ug/L	100
20) cis-1,2-Dichloroethene	4.671	96	2203	0.685	ug/L	74
22) 2-Butanone	4.706	43	23072	7.973	ug/L	93
46) Tetrachloroethene	8.552	164	12404	5.332	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112022\
Data File : VU051981.D
Acq On : 20 Nov 2022 18:14
Operator : JC/MD
Sample : N5710-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0KY4

Quant Time: Nov 21 01:03:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM111522WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 18 00:53:49 2022
Response via : Initial Calibration

