

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112024\
 Data File : VU061872.D
 Acq On : 20 Nov 2024 17:41
 Operator : MD/SY
 Sample : P4847-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A0AT7

Quant Time: Nov 21 01:30:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 21 01:18:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	194012	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	215688	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	79294	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	54285	5.704	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.000%
7) Chloroethane-d5	1.907	69	44339	5.267	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.400%
11) 1,1-Dichloroethene-d2	2.560	65	23970	6.305	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	126.000%#
20) 2-Butanone-d5	4.624	46	117363	39.272	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.540%
24) Chloroform-d	5.052	84	118254	5.093	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
26) 1,2-Dichloroethane-d4	5.692	65	52895	4.427	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.600%
32) Benzene-d6	5.718	84	203526	4.843	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
36) 1,2-Dichloropropane-d6	6.682	67	64653	4.528	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.600%
41) Toluene-d8	7.891	98	176635	5.040	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
43) trans-1,3-Dichloroprop...	8.174	79	21171	4.005	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.000%
46) 2-Hexanone-d5	8.627	63	76273	31.611	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	63.220%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	46533	3.261	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	65.200%
66) 1,2-Dichlorobenzene-d4	12.184	152	58213	4.965	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.200%
Target Compounds						
13) Acetone	2.637	43	138400	67.077	ug/L	99
25) Chloroform	5.075	83	360248	12.863	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112024\
Data File : VU061872.D
Acq On : 20 Nov 2024 17:41
Operator : MD/SY
Sample : P4847-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A0AT7

Quant Time: Nov 21 01:30:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 21 01:18:14 2024
Response via : Initial Calibration

