

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112024\
 Data File : VU061874.D
 Acq On : 20 Nov 2024 18:31
 Operator : MD/SY
 Sample : P4827-03RE
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H46W3RE

Quant Time: Nov 21 01:31:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 21 01:18:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	167319	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	156777	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	76999	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	55911	6.812	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 136.200%#		
7) Chloroethane-d5	1.904	69	44406	6.116	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 122.400%		
11) 1,1-Dichloroethene-d2	2.560	65	23546	7.181	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 143.600%#		
20) 2-Butanone-d5	4.618	46	108190	41.979	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 83.960%		
24) Chloroform-d	5.052	84	102548	5.122	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 102.400%		
26) 1,2-Dichloroethane-d4	5.692	65	53774	5.218	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 104.400%		
32) Benzene-d6	5.718	84	203178	6.652	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 133.000%#		
36) 1,2-Dichloropropane-d6	6.682	67	63548	6.123	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 122.400%		
41) Toluene-d8	7.891	98	175485	6.888	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 137.800%#		
43) trans-1,3-Dichloroprop...	8.174	79	19098	4.970	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 99.400%		
46) 2-Hexanone-d5	8.624	63	72589	41.389	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 82.780%		
56) 1,1,2,2-Tetrachloroeth...	10.746	84	54696	5.274	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 105.400%		
66) 1,2-Dichlorobenzene-d4	12.187	152	62928	5.527	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 110.600%		
Target Compounds						Qvalue
13) Acetone	2.628	43	26718	15.015	ug/L	97
17) Methyl tert-butyl Ether	3.354	73	48114	1.673	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112024\
Data File : VU061874.D
Acq On : 20 Nov 2024 18:31
Operator : MD/SY
Sample : P4827-03RE
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H46W3RE

Quant Time: Nov 21 01:31:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 21 01:18:14 2024
Response via : Initial Calibration

