

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112024\
 Data File : VU061897.D
 Acq On : 21 Nov 2024 05:18
 Operator : MD/SY
 Sample : P4875-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2A15

Quant Time: Nov 21 07:35:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 21 03:29:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	170117	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	177193	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	67229	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	51790	6.206	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	124.200%
7) Chloroethane-d5	1.907	69	41906	5.677	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.600%
11) 1,1-Dichloroethene-d2	2.560	65	22089	6.626	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	132.600%#
20) 2-Butanone-d5	4.612	46	110561	42.193	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.380%
24) Chloroform-d	5.052	84	98693	4.848	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
26) 1,2-Dichloroethane-d4	5.692	65	52503	5.011	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
32) Benzene-d6	5.718	84	197404	5.718	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.400%
36) 1,2-Dichloropropane-d6	6.682	67	61806	5.269	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.400%
41) Toluene-d8	7.891	98	191538	6.652	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	133.000%#
43) trans-1,3-Dichloroprop...	8.174	79	19842	4.569	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.624	63	76343	38.514	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.020%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	45738	3.902	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	78.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	53085	5.340	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112024\
Data File : VU061897.D
Acq On : 21 Nov 2024 05:18
Operator : MD/SY
Sample : P4875-06
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2A15

Quant Time: Nov 21 07:35:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 21 03:29:06 2024
Response via : Initial Calibration

