

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112024\
 Data File : VU061899.D
 Acq On : 21 Nov 2024 06:07
 Operator : MD/SY
 Sample : P4875-08
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2A22

Quant Time: Nov 21 11:56:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 21 03:29:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	176759	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	169349	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	72166	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	52005	5.998	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.000%
7) Chloroethane-d5	1.907	69	41325	5.388	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.800%
11) 1,1-Dichloroethene-d2	2.560	65	22489	6.492	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	129.800%#
20) 2-Butanone-d5	4.615	46	104219	38.278	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.560%
24) Chloroform-d	5.052	84	98872	4.674	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
26) 1,2-Dichloroethane-d4	5.692	65	52248	4.800	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
32) Benzene-d6	5.717	84	195629	5.929	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.600%
36) 1,2-Dichloropropane-d6	6.682	67	61949	5.525	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.600%
41) Toluene-d8	7.891	98	187348	6.808	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	136.200%#
43) trans-1,3-Dichloroprop...	8.174	79	22149	5.336	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.800%
46) 2-Hexanone-d5	8.624	63	80629	42.561	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.120%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	44899	4.008	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	52316	4.903	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.000%
Target Compounds						
16) Methylene chloride	3.036	84	17172	1.206	ug/L	94
25) Chloroform	5.078	83	15092	0.591	ug/L	91
42) Toluene	7.962	91	117485	2.189	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112024\
Data File : VU061899.D
Acq On : 21 Nov 2024 06:07
Operator : MD/SY
Sample : P4875-08
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2A22

Quant Time: Nov 21 11:56:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 21 03:29:06 2024
Response via : Initial Calibration

