

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112224\
 Data File : VU061959.D
 Acq On : 22 Nov 2024 09:58
 Operator : MD/SY
 Sample : P4885-17
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 22 22:42:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 22 22:40:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	142132	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	132536	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	58348	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	27155	3.895	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.800%
7) Chloroethane-d5	1.907	69	24594	3.988	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.800%
11) 1,1-Dichloroethene-d2	2.560	65	12371	4.441	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.800%
20) 2-Butanone-d5	4.618	46	100334	45.829	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.660%
24) Chloroform-d	5.055	84	76092	4.474	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
26) 1,2-Dichloroethane-d4	5.692	65	39911	4.560	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.200%
32) Benzene-d6	5.721	84	127864	4.952	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	6.682	67	46880	5.343	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.800%
41) Toluene-d8	7.891	98	103902	4.824	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
43) trans-1,3-Dichloroprop...	8.174	79	15693	4.831	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.600%
46) 2-Hexanone-d5	8.627	63	66869	45.101	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.200%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	40330	4.600	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	44122	5.114	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112224\
Data File : VU061959.D
Acq On : 22 Nov 2024 09:58
Operator : MD/SY
Sample : P4885-17
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 22 22:42:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 22 22:40:46 2024
Response via : Initial Calibration

