

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112224\
 Data File : VU061974.D
 Acq On : 22 Nov 2024 16:04
 Operator : MD/SY
 Sample : P4903-05DL 10X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG401DL

Quant Time: Nov 22 22:46:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 22 22:40:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	132313	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	123580	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	56471	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	32513	5.010	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.200%
7) Chloroethane-d5	1.898	69	19613	3.416	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.400%
11) 1,1-Dichloroethene-d2	2.557	65	15308	5.904	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	118.000%
20) 2-Butanone-d5	4.621	46	78534	38.534	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.060%
24) Chloroform-d	5.052	84	68500	4.326	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
26) 1,2-Dichloroethane-d4	5.692	65	36408	4.468	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
32) Benzene-d6	5.718	84	136809	5.682	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.600%
36) 1,2-Dichloropropane-d6	6.679	67	43053	5.262	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.200%
41) Toluene-d8	7.891	98	118770	5.914	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.200%
43) trans-1,3-Dichloroprop...	8.174	79	14746	4.869	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.400%
46) 2-Hexanone-d5	8.631	63	54822	39.656	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.320%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	32905	4.025	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	40366	4.834	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%
Target Compounds					Qvalue	
34) Trichloroethene	6.534	95	16671	1.674	ug/L	97
47) Tetrachloroethene	8.544	164	60146	8.302	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112224\
Data File : VU061974.D
Acq On : 22 Nov 2024 16:04
Operator : MD/SY
Sample : P4903-05DL 10X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG401DL

Quant Time: Nov 22 22:46:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 22 22:40:46 2024
Response via : Initial Calibration

