

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112224\
 Data File : VU061985.D
 Acq On : 22 Nov 2024 21:22
 Operator : MD/SY
 Sample : P4904-01DL 5X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG411DL

Quant Time: Nov 23 00:15:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 23 00:14:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	86524	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	78943	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	34153	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	28641	6.748	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	= 135.000%#	
7) Chloroethane-d5	1.907	69	23064	6.143	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	= 122.800%	
11) 1,1-Dichloroethene-d2	2.560	65	12668	7.471	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	= 149.400%#	
20) 2-Butanone-d5	4.621	46	70620	52.988	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	= 105.980%	
24) Chloroform-d	5.052	84	63218	6.106	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 122.200%	
26) 1,2-Dichloroethane-d4	5.692	65	33400	6.268	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 125.400%	
32) Benzene-d6	5.717	84	117139	7.616	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 152.400%#	
36) 1,2-Dichloropropane-d6	6.682	67	38334	7.335	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	= 146.600%#	
41) Toluene-d8	7.891	98	99191	7.732	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 154.600%#	
43) trans-1,3-Dichloroprop...	8.177	79	12689	6.558	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	= 131.200%#	
46) 2-Hexanone-d5	8.627	63	48742	55.194	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	= 110.380%	
56) 1,1,2,2-Tetrachloroeth...	10.746	84	29851	5.716	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	= 114.400%	
66) 1,2-Dichlorobenzene-d4	12.187	152	33989	6.730	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	= 134.600%#	
Target Compounds						Qvalue
9) Trichlorofluoromethane	2.132	101	80602	7.949	ug/L	99
12) 1,1-Dichloroethene	2.573	96	1435	0.247	ug/L #	1
16) Methylene chloride	3.039	84	5083	0.730	ug/L	93
17) Methyl tert-butyl Ether	3.354	73	11067	0.744	ug/L #	85
19) 1,1-Dichloroethane	3.862	63	9031	0.756	ug/L	100
22) cis-1,2-Dichloroethene	4.663	96	1183	0.171	ug/L	97
29) 1,1,1-Trichloroethane	5.306	97	4763	0.523	ug/L	95
34) Trichloroethene	6.534	95	10355	1.627	ug/L	95
47) Tetrachloroethene	8.544	164	4617	0.998	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112224\
Data File : VU061985.D
Acq On : 22 Nov 2024 21:22
Operator : MD/SY
Sample : P4904-01DL 5X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG411DL

Quant Time: Nov 23 00:15:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 23 00:14:12 2024
Response via : Initial Calibration

