

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112324\
 Data File : VU062013.D
 Acq On : 23 Nov 2024 17:00
 Operator : MD/SY
 Sample : P4904-20
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 25 01:27:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 25 01:25:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	105998	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	100482	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	41881	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	18626	3.582	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.600%
7) Chloroethane-d5	1.908	69	16561	3.601	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.000%
11) 1,1-Dichloroethene-d2	2.560	65	8677	4.177	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.600%
20) 2-Butanone-d5	4.621	46	75562	46.280	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.560%
24) Chloroform-d	5.052	84	58049	4.576	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.692	65	31462	4.820	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.718	84	95060	4.856	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
36) 1,2-Dichloropropane-d6	6.682	67	33607	5.052	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.000%
41) Toluene-d8	7.891	98	75408	4.618	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.400%
43) trans-1,3-Dichloroprop...	8.174	79	12153	4.935	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.600%
46) 2-Hexanone-d5	8.627	63	51431	45.755	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.500%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	31008	4.665	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	34452	5.563	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112324\
Data File : VU062013.D
Acq On : 23 Nov 2024 17:00
Operator : MD/SY
Sample : P4904-20
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 25 01:27:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 25 01:25:39 2024
Response via : Initial Calibration

