

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112324\
 Data File : VU062008.D
 Acq On : 23 Nov 2024 14:51
 Operator : MD/SY
 Sample : VU1123WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK098

Quant Time: Nov 25 01:26:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 25 01:25:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	117430	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	110676	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	48114	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	20087	3.487	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.800%
7) Chloroethane-d5	1.907	69	17063	3.349	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	67.000%
11) 1,1-Dichloroethene-d2	2.560	65	9029	3.923	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.400%
20) 2-Butanone-d5	4.621	46	75529	41.756	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.520%
24) Chloroform-d	5.052	84	59312	4.221	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.400%
26) 1,2-Dichloroethane-d4	5.692	65	31938	4.416	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
32) Benzene-d6	5.718	84	98444	4.565	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.682	67	34571	4.718	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.400%
41) Toluene-d8	7.891	98	79623	4.427	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.600%
43) trans-1,3-Dichloroprop...	8.174	79	12327	4.544	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.800%
46) 2-Hexanone-d5	8.627	63	52201	42.162	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.320%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	31991	4.369	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	36058	5.068	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112324\
Data File : VU062008.D
Acq On : 23 Nov 2024 14:51
Operator : MD/SY
Sample : VU1123WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK098

Quant Time: Nov 25 01:26:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 25 01:25:39 2024
Response via : Initial Calibration

