

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112524\
 Data File : VU062028.D
 Acq On : 25 Nov 2024 17:23
 Operator : MD/SY
 Sample : P4996-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHF74

Quant Time: Nov 26 03:47:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 25 01:25:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	130893	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	119477	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	54670	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	0.000	65	0	0.000	ug/L	
Spiked Amount	5.000	Range	40 - 130	Recovery	=	0.000%#
7) Chloroethane-d5	1.907	69	22337	3.933	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.600%
11) 1,1-Dichloroethene-d2	0.000	65	0	0.000	ug/L	
Spiked Amount	5.000	Range	60 - 125	Recovery	=	0.000%#
20) 2-Butanone-d5	4.618	46	62087	30.794	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	61.580%
24) Chloroform-d	5.055	84	62704	4.003	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.000%
26) 1,2-Dichloroethane-d4	5.695	65	32588	4.043	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.800%
32) Benzene-d6	5.717	84	119322	5.126	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.600%
36) 1,2-Dichloropropane-d6	6.682	67	37579	4.751	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.910	98	420	0.022	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	0.400%#
43) trans-1,3-Dichloroprop...	8.190	79	312	0.107	ug/L	0.02
Spiked Amount	5.000	Range	55 - 130	Recovery	=	2.200%#
46) 2-Hexanone-d5	8.627	63	43147	32.282	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	64.560%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	28143	3.561	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	71.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	37829	4.680	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112524\
Data File : VU062028.D
Acq On : 25 Nov 2024 17:23
Operator : MD/SY
Sample : P4996-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHF74

Quant Time: Nov 26 03:47:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 25 01:25:39 2024
Response via : Initial Calibration

