

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112624\
 Data File : VU062038.D
 Acq On : 26 Nov 2024 11:37
 Operator : MD/SY
 Sample : P4996-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHF74

Quant Time: Nov 27 04:17:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 25 01:25:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	136087	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	127688	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	58062	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	0.000	65	0	0.000	ug/L	
Spiked Amount	5.000	Range	40 - 130	Recovery	=	0.000%#
7) Chloroethane-d5	1.907	69	21890	3.707	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.200%
11) 1,1-Dichloroethene-d2	2.560	65	896	0.336	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	6.800%#
20) 2-Butanone-d5	4.618	46	58699	28.003	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	56.000%
24) Chloroform-d	5.052	84	59093	3.629	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.600%
26) 1,2-Dichloroethane-d4	5.695	65	31816	3.796	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.000%
32) Benzene-d6	5.718	84	113608	4.567	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.682	67	35903	4.247	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.000%
41) Toluene-d8	7.897	98	10498	0.506	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	10.200%#
43) trans-1,3-Dichloroprop...	8.180	79	3438	1.099	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	22.000%#
46) 2-Hexanone-d5	8.627	63	40581	28.410	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	56.820%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26197	3.101	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	62.000%#
66) 1,2-Dichlorobenzene-d4	12.187	152	37514	4.369	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112624\
Data File : VU062038.D
Acq On : 26 Nov 2024 11:37
Operator : MD/SY
Sample : P4996-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHF74

Quant Time: Nov 27 04:17:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 25 01:25:39 2024
Response via : Initial Calibration

