

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120224\
 Data File : VU062068.D
 Acq On : 02 Dec 2024 11:51
 Operator : MD/SY
 Sample : VU1202WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK102

Quant Time: Dec 03 00:24:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 25 01:25:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	114554	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	106011	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	47509	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	19936	3.548	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.000%
7) Chloroethane-d5	1.911	69	17443	3.509	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.200%
11) 1,1-Dichloroethene-d2	2.560	65	9065	4.038	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.800%
20) 2-Butanone-d5	4.637	46	63642	36.068	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.140%
24) Chloroform-d	5.052	84	53305	3.889	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.800%
26) 1,2-Dichloroethane-d4	5.692	65	28380	4.023	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
32) Benzene-d6	5.718	84	88875	4.303	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
36) 1,2-Dichloropropane-d6	6.679	67	30907	4.404	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.000%
41) Toluene-d8	7.891	98	75474	4.381	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
43) trans-1,3-Dichloroprop...	8.174	79	11616	4.471	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.400%
46) 2-Hexanone-d5	8.631	63	42825	36.111	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.220%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26179	3.733	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	30304	4.314	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120224\
Data File : VU062068.D
Acq On : 02 Dec 2024 11:51
Operator : MD/SY
Sample : VU1202WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK102

Quant Time: Dec 03 00:24:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 25 01:25:46 2024
Response via : Initial Calibration

