

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120224\
 Data File : VU062071.D
 Acq On : 02 Dec 2024 13:27
 Operator : MD/SY
 Sample : P4996-09
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Dec 03 00:25:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 03 00:21:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	119434	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	113904	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	52081	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	21431	3.658	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.200%
7) Chloroethane-d5	1.911	69	18758	3.619	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.400%
11) 1,1-Dichloroethene-d2	2.563	65	9346	3.993	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.800%
20) 2-Butanone-d5	4.634	46	65401	35.550	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	71.100%
24) Chloroform-d	5.052	84	56332	3.941	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.800%
26) 1,2-Dichloroethane-d4	5.695	65	30664	4.169	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.400%
32) Benzene-d6	5.717	84	96903	4.366	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.400%
36) 1,2-Dichloropropane-d6	6.682	67	33749	4.475	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.600%
41) Toluene-d8	7.891	98	80899	4.371	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	8.177	79	12764	4.572	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.631	63	46016	36.113	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.220%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	27953	3.710	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.200%
66) 1,2-Dichlorobenzene-d4	12.183	152	34903	4.532	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120224\
Data File : VU062071.D
Acq On : 02 Dec 2024 13:27
Operator : MD/SY
Sample : P4996-09
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Dec 03 00:25:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 03 00:21:51 2024
Response via : Initial Calibration

