

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120424\
 Data File : VU062088.D
 Acq On : 04 Dec 2024 15:20
 Operator : MD/SY
 Sample : VSTDCCC0.5EC
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5107

Quant Time: Dec 05 01:41:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 05 01:30:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	16829	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	23745	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	8606	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	5892	0.495	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 100.000%			
4) 1,2-Dichloroethane-d4	5.685	65	6258	0.509	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 102.000%			
7) 1,2-Dichloropropane-d6	6.687	67	6499	0.481	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 96.000%			
8) Toluene-d8	7.887	98	17795	0.490	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 98.000%			
10) 1,1,2,2-Tetrachloroeth...	10.745	84	5456	0.481	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 96.000%			
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.597	62	7130	0.466	ug/L #	69
6) Trichloroethene	6.544	95	6596	0.472	ug/L	99
9) 1,2-Dibromoethane	8.924	107	4109	0.462	ug/L	98
12) 1,2,3-Trichloropropane	10.814	75	3900	0.428	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.990	75	771	0.442	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120424\
Data File : VU062088.D
Acq On : 04 Dec 2024 15:20
Operator : MD/SY
Sample : VSTDCCC0.5EC
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5107

Quant Time: Dec 05 01:41:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 05 01:30:27 2024
Response via : Initial Calibration

