

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120524\
 Data File : VU062091.D
 Acq On : 05 Dec 2024 10:48
 Operator : MD/SY
 Sample : VU1205MBL01
 Misc : 5.00g/5mL/100uL/5ML/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Dec 06 00:44:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112724WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 06 00:44:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	176906	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	163867	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	76267	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	56043	45.723	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.440%
7) Chloroethane-d5	1.888	69	31011	35.105	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	70.200%
11) 1,1-Dichloroethene-d2	2.541	63	84248	36.252	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.500%
21) 2-Butanone-d5	4.627	46	75368	90.486	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.490%
24) Chloroform-d	5.052	84	119366	46.660	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.320%
26) 1,2-Dichloroethane-d4	5.692	65	80023	49.032	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.060%
32) Benzene-d6	5.714	84	233934	49.923	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.840%
36) 1,2-Dichloropropane-d6	6.679	67	73587	49.971	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.940%
41) Toluene-d8	7.891	98	215367	49.647	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.300%
43) trans-1,3-Dichloroprop...	8.174	79	36119	49.079	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.160%
47) 2-Hexanone-d5	8.650	63	53049	95.633	ug/L	0.03
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.630%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	104196	48.622	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.240%
66) 1,2-Dichlorobenzene-d4	12.187	152	77988	51.979	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120524\
Data File : VU062091.D
Acq On : 05 Dec 2024 10:48
Operator : MD/SY
Sample : VU1205MBL01
Misc : 5.00g/5mL/100ul/5ML/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Dec 06 00:44:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112724WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 06 00:44:17 2024
Response via : Initial Calibration

