

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120624\
 Data File : VU062105.D
 Acq On : 05 Dec 2024 17:58
 Operator : MD/SY
 Sample : P5104-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHJ64

Quant Time: Dec 06 02:29:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 06 02:27:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	104161	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	98095	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	46439	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	21818	4.270	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.400%
7) Chloroethane-d5	1.907	69	19048	4.214	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.200%
11) 1,1-Dichloroethene-d2	2.560	65	9951	4.875	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.400%
20) 2-Butanone-d5	4.628	46	71613	44.635	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.260%
24) Chloroform-d	5.052	84	61468	4.931	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
26) 1,2-Dichloroethane-d4	5.692	65	33977	5.297	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.000%
32) Benzene-d6	5.718	84	102758	5.377	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.600%
36) 1,2-Dichloropropane-d6	6.682	67	35013	5.391	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.800%
41) Toluene-d8	7.891	98	85157	5.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
43) trans-1,3-Dichloroprop...	8.174	79	13664	5.683	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	113.600%
46) 2-Hexanone-d5	8.627	63	50347	45.880	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.760%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	30372	4.680	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	38769	5.646	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120624\
Data File : VU062105.D
Acq On : 05 Dec 2024 17:58
Operator : MD/SY
Sample : P5104-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHJ64

Quant Time: Dec 06 02:29:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 06 02:27:00 2024
Response via : Initial Calibration

