

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120724\
 Data File : VU062141.D
 Acq On : 06 Dec 2024 18:07
 Operator : MD/SY
 Sample : P5107-03
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHJ80

Quant Time: Dec 06 23:17:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 06 02:27:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	102162	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	98747	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	45667	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	16780	3.348	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.000%
7) Chloroethane-d5	1.908	69	15094	3.405	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.000%
11) 1,1-Dichloroethene-d2	2.560	65	8222	4.107	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.200%
20) 2-Butanone-d5	4.624	46	61393	39.013	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.020%
24) Chloroform-d	5.052	84	52979	4.334	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
26) 1,2-Dichloroethane-d4	5.692	65	29158	4.634	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	5.718	84	87491	4.547	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
36) 1,2-Dichloropropane-d6	6.682	67	30226	4.624	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.891	98	73519	4.582	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.600%
43) trans-1,3-Dichloroprop...	8.174	79	11282	4.662	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.200%
46) 2-Hexanone-d5	8.628	63	41927	37.955	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.900%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26712	4.089	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.800%
66) 1,2-Dichlorobenzene-d4	12.184	152	32799	4.857	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120724\
Data File : VU062141.D
Acq On : 06 Dec 2024 18:07
Operator : MD/SY
Sample : P5107-03
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHJ80

Quant Time: Dec 06 23:17:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 06 02:27:00 2024
Response via : Initial Calibration

