

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121024\
 Data File : VU062214.D
 Acq On : 10 Dec 2024 17:50
 Operator : MD/SY
 Sample : P5107-19
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHJ98

Quant Time: Dec 11 03:46:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 11 03:43:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.241	114	97294	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	94389	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	43884	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	27771	4.232	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.600%
7) Chloroethane-d5	1.907	69	20296	4.348	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.000%
11) 1,1-Dichloroethene-d2	2.554	65	14586	4.672	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.400%
20) 2-Butanone-d5	4.627	46	73854	58.721	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.440%
24) Chloroform-d	5.052	84	68110	4.795	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
26) 1,2-Dichloroethane-d4	5.692	65	37782	5.410	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.200%
32) Benzene-d6	5.714	84	125103	4.602	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
36) 1,2-Dichloropropane-d6	6.682	67	37570	4.830	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	7.891	98	111552	4.344	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	8.174	79	15666	4.576	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.600%
46) 2-Hexanone-d5	8.634	63	63030	61.021	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	122.040%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	32556	5.379	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	42770	5.289	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121024\
Data File : VU062214.D
Acq On : 10 Dec 2024 17:50
Operator : MD/SY
Sample : P5107-19
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHJ98

Quant Time: Dec 11 03:46:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 11 03:43:25 2024
Response via : Initial Calibration

