

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121024\
 Data File : VU062217.D
 Acq On : 10 Dec 2024 19:04
 Operator : MD/SY
 Sample : P5148-04
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH7S9

Quant Time: Dec 11 03:47:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 11 03:43:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	99391	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	94057	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	44037	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	26617	3.970	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.400%
7) Chloroethane-d5	1.907	69	20177	4.231	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.600%
11) 1,1-Dichloroethene-d2	2.550	65	13697	4.295	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.800%
20) 2-Butanone-d5	4.634	46	70380	54.778	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.560%
24) Chloroform-d	5.052	84	66380	4.575	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.692	65	36722	5.147	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.000%
32) Benzene-d6	5.714	84	121714	4.493	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
36) 1,2-Dichloropropane-d6	6.679	67	35765	4.614	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.200%
41) Toluene-d8	7.888	98	107424	4.198	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	8.174	79	14846	4.351	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.634	63	57577	55.938	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.880%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	30807	5.108	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	41276	5.086	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121024\
Data File : VU062217.D
Acq On : 10 Dec 2024 19:04
Operator : MD/SY
Sample : P5148-04
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH7S9

Quant Time: Dec 11 03:47:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 11 03:43:25 2024
Response via : Initial Calibration

