

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121024\
 Data File : VU062225.D
 Acq On : 10 Dec 2024 22:21
 Operator : MD/SY
 Sample : P5247-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH7T6

Quant Time: Dec 11 03:50:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 11 03:43:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	99073	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	93799	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	43682	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	27790	4.158	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.200%
7) Chloroethane-d5	1.904	69	20660	4.346	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.000%
11) 1,1-Dichloroethene-d2	2.554	65	13559	4.265	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.200%
20) 2-Butanone-d5	4.628	46	65173	50.888	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.780%
24) Chloroform-d	5.052	84	66982	4.631	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	5.692	65	36856	5.182	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.600%
32) Benzene-d6	5.714	84	125846	4.658	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
36) 1,2-Dichloropropane-d6	6.679	67	37886	4.901	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.000%
41) Toluene-d8	7.891	98	110798	4.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	8.174	79	15107	4.440	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.800%
46) 2-Hexanone-d5	8.631	63	52880	51.516	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.040%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	29458	4.898	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	41235	5.123	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121024\
Data File : VU062225.D
Acq On : 10 Dec 2024 22:21
Operator : MD/SY
Sample : P5247-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH7T6

Quant Time: Dec 11 03:50:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 11 03:43:25 2024
Response via : Initial Calibration

