

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121121\
 Data File : VU046287.D
 Acq On : 11 Dec 2021 15:27
 Operator : SY/MD
 Sample : M4943-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Dec 12 01:38:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	173446	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	179643	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	93738	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	59009	41.301	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.600%
7) Chloroethane-d5	1.916	69	48693	44.388	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.780%
11) 1,1-Dichloroethene-d2	2.572	63	84933	33.114	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.220%
21) 2-Butanone-d5	4.636	46	135214	119.048	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.050%
24) Chloroform-d	5.067	84	125042	52.408	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.820%
26) 1,2-Dichloroethane-d4	5.706	65	87294	54.516	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.040%
32) Benzene-d6	5.732	84	251249	48.797	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.600%
36) 1,2-Dichloropropane-d6	6.694	67	83798	52.514	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.020%
41) Toluene-d8	7.899	98	218182	46.540	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.080%
43) trans-1,3-Dichloroprop...	8.182	79	39843	51.723	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.440%
47) 2-Hexanone-d5	8.636	63	97038	128.256	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	128.260%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	137993	57.022	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	114.040%
66) 1,2-Dichlorobenzene-d4	12.195	152	95706	53.016	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.040%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121121\
Data File : VU046287.D
Acq On : 11 Dec 2021 15:27
Operator : SY/MD
Sample : M4943-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Dec 12 01:38:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

