

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121124\
 Data File : VU062270.D
 Acq On : 12 Dec 2024 04:26
 Operator : MD/SY
 Sample : P5251-18
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7J44

Quant Time: Dec 12 05:59:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 12 01:45:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	78979	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	74962	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	33740	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	22016	4.133	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.600%
7) Chloroethane-d5	1.908	69	18996	5.013	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.200%
11) 1,1-Dichloroethene-d2	2.560	65	10949	4.320	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.400%
20) 2-Butanone-d5	4.624	46	63380	62.079	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.160%
24) Chloroform-d	5.052	84	61528	5.337	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.800%
26) 1,2-Dichloroethane-d4	5.695	65	33718	5.947	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.000%
32) Benzene-d6	5.718	84	104868	4.857	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
36) 1,2-Dichloropropane-d6	6.679	67	32941	5.332	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.600%
41) Toluene-d8	7.891	98	91564	4.489	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.800%
43) trans-1,3-Dichloroprop...	8.174	79	12767	4.695	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.000%
46) 2-Hexanone-d5	8.627	63	47777	58.241	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.480%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	27029	5.623	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	35209	5.663	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.200%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.573	96	1798	0.337	ug/L	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121124\
Data File : VU062270.D
Acq On : 12 Dec 2024 04:26
Operator : MD/SY
Sample : P5251-18
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7J44

Quant Time: Dec 12 05:59:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 12 01:45:00 2024
Response via : Initial Calibration

