

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121124\
 Data File : VU062233.D
 Acq On : 11 Dec 2024 12:03
 Operator : MD/SY
 Sample : VIBLK182
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK182

Quant Time: Dec 12 00:25:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 12 00:24:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	71274	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	64336	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	28640	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	11709	2.436	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	48.800%
7) Chloroethane-d5	1.908	69	12385	3.622	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.400%
11) 1,1-Dichloroethene-d2	2.560	65	5260	2.300	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	46.000%#
20) 2-Butanone-d5	4.634	46	58571	63.571	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.140%
24) Chloroform-d	5.052	84	45503	4.373	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.400%
26) 1,2-Dichloroethane-d4	5.695	65	27355	5.347	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.000%
32) Benzene-d6	5.718	84	65276	3.523	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	70.400%
36) 1,2-Dichloropropane-d6	6.682	67	23612	4.454	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.000%
41) Toluene-d8	7.891	98	47669	2.723	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	54.400%#
43) trans-1,3-Dichloroprop...	8.178	79	7189	3.081	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.600%
46) 2-Hexanone-d5	8.631	63	37981	53.947	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.900%
56) 1,1,2,2-Tetrachloroeth...	10.747	84	24156	5.855	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	117.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	23044	4.366	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121124\
Data File : VU062233.D
Acq On : 11 Dec 2024 12:03
Operator : MD/SY
Sample : VIBLK182
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK182

Quant Time: Dec 12 00:25:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 12 00:24:05 2024
Response via : Initial Calibration

