

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121224\
 Data File : VU062275.D
 Acq On : 12 Dec 2024 09:56
 Operator : MD/SY
 Sample : VU1212WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK111

Quant Time: Dec 13 00:14:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 13 00:13:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	91439	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	85705	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	39413	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	23338	3.784	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.600%
7) Chloroethane-d5	1.907	69	20203	4.605	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.200%
11) 1,1-Dichloroethene-d2	2.557	65	11775	4.013	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.200%
20) 2-Butanone-d5	4.631	46	59830	50.617	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.240%
24) Chloroform-d	5.052	84	61306	4.593	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.800%
26) 1,2-Dichloroethane-d4	5.692	65	32429	4.941	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.800%
32) Benzene-d6	5.714	84	109710	4.445	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	6.679	67	32692	4.629	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.600%
41) Toluene-d8	7.891	98	96767	4.150	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.000%
43) trans-1,3-Dichloroprop...	8.174	79	13512	4.346	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.627	63	45841	48.876	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.760%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26469	4.816	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.400%
66) 1,2-Dichlorobenzene-d4	12.183	152	35959	4.951	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121224\
Data File : VU062275.D
Acq On : 12 Dec 2024 09:56
Operator : MD/SY
Sample : VU1212WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK111

Quant Time: Dec 13 00:14:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 13 00:13:55 2024
Response via : Initial Calibration

