

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121224\
 Data File : VU062286.D
 Acq On : 12 Dec 2024 14:28
 Operator : MD/SY
 Sample : P5253-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7J10

Quant Time: Dec 13 00:17:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 13 00:13:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	81757	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	75287	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	33217	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	21313	3.865	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.200%
7) Chloroethane-d5	1.911	69	18834	4.802	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.000%
11) 1,1-Dichloroethene-d2	2.560	65	11129	4.242	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.800%
20) 2-Butanone-d5	4.628	46	60088	56.855	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.720%
24) Chloroform-d	5.052	84	58564	4.907	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	5.692	65	33544	5.716	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.400%
32) Benzene-d6	5.718	84	102740	4.738	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
36) 1,2-Dichloropropane-d6	6.682	67	31565	5.088	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.800%
41) Toluene-d8	7.891	98	91221	4.453	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.000%
43) trans-1,3-Dichloroprop...	8.174	79	13360	4.892	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.800%
46) 2-Hexanone-d5	8.627	63	46120	55.979	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.960%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	25968	5.379	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.600%
66) 1,2-Dichlorobenzene-d4	12.184	152	34154	5.580	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121224\
Data File : VU062286.D
Acq On : 12 Dec 2024 14:28
Operator : MD/SY
Sample : P5253-06
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7J10

Quant Time: Dec 13 00:17:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 13 00:13:55 2024
Response via : Initial Calibration

