

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121324\
 Data File : VU062303.D
 Acq On : 13 Dec 2024 11:19
 Operator : MD/SY
 Sample : P5228-07DL 5X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E29S9DL

Quant Time: Dec 14 03:47:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Dec 14 03:45:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	90947	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	87036	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	40567	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	20386	3.323	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.400%
7) Chloroethane-d5	1.911	69	18061	4.139	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.800%
11) 1,1-Dichloroethene-d2	2.557	65	10499	3.598	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.000%
20) 2-Butanone-d5	4.634	46	61648	52.437	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.880%
24) Chloroform-d	5.049	84	55594	4.187	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
26) 1,2-Dichloroethane-d4	5.692	65	30857	4.726	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.600%
32) Benzene-d6	5.717	84	98803	3.941	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.800%
36) 1,2-Dichloropropane-d6	6.679	67	31132	4.340	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.800%
41) Toluene-d8	7.888	98	85995	3.631	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.600%
43) trans-1,3-Dichloroprop...	8.174	79	12368	3.918	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.400%
46) 2-Hexanone-d5	8.631	63	48993	51.438	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.880%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26771	4.797	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.000%
66) 1,2-Dichlorobenzene-d4	12.183	152	34273	4.585	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.600%
Target Compounds						Qvalue
47) Tetrachloroethene	8.544	164	67860	11.503	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121324\
Data File : VU062303.D
Acq On : 13 Dec 2024 11:19
Operator : MD/SY
Sample : P5228-07DL 5X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E29S9DL

Quant Time: Dec 14 03:47:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Dec 14 03:45:56 2024
Response via : Initial Calibration

