

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121324\
 Data File : VU062307.D
 Acq On : 13 Dec 2024 12:56
 Operator : MD/SY
 Sample : VIBLK186
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK186

Quant Time: Dec 14 03:48:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Dec 14 03:45:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	89833	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	86519	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	38873	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	20860	3.443	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.800%
7) Chloroethane-d5	1.911	69	18614	4.319	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.400%
11) 1,1-Dichloroethene-d2	2.560	65	10929	3.791	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.800%
20) 2-Butanone-d5	4.631	46	60403	52.015	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.040%
24) Chloroform-d	5.052	84	58372	4.451	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
26) 1,2-Dichloroethane-d4	5.692	65	31010	4.809	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.200%
32) Benzene-d6	5.717	84	101517	4.074	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.400%
36) 1,2-Dichloropropane-d6	6.679	67	31314	4.392	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.800%
41) Toluene-d8	7.891	98	88918	3.777	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.600%
43) trans-1,3-Dichloroprop...	8.174	79	12361	3.939	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.800%
46) 2-Hexanone-d5	8.627	63	46704	49.328	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.660%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	25689	4.630	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.600%
66) 1,2-Dichlorobenzene-d4	12.183	152	34106	4.761	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121324\
Data File : VU062307.D
Acq On : 13 Dec 2024 12:56
Operator : MD/SY
Sample : VIBLK186
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK186

Quant Time: Dec 14 03:48:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Dec 14 03:45:56 2024
Response via : Initial Calibration

