

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121423\
 Data File : VU056837.D
 Acq On : 14 Dec 2023 14:29
 Operator : MD/SY
 Sample : 05806-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E0LR3

Quant Time: Dec 15 00:52:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 15 00:49:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	240432	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	246776	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	98546	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	73783	3.580	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.600%
7) Chloroethane-d5	1.911	69	65609	4.254	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.000%
11) 1,1-Dichloroethene-d2	2.563	65	35091	3.980	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.600%
20) 2-Butanone-d5	4.621	46	226254	55.591	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.180%
24) Chloroform-d	5.055	84	187971	4.817	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
26) 1,2-Dichloroethane-d4	5.698	65	93820	5.080	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
32) Benzene-d6	5.721	84	343528	4.488	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
36) 1,2-Dichloropropane-d6	6.685	67	113740	4.748	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.894	98	274824	4.017	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
43) trans-1,3-Dichloroprop...	8.177	79	38161	4.107	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.200%
46) 2-Hexanone-d5	8.630	63	162403	46.346	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.700%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	86580	4.995	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	97672	5.042	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.800%
Target Compounds					Qvalue	
13) Acetone	2.634	43	12978	5.232	ug/L	65
42) Toluene	7.968	91	24053	0.268	ug/L	91
53) m,p-Xylene	9.695	106	9522	0.258	ug/L	80


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121423\
Data File : VU056837.D
Acq On : 14 Dec 2023 14:29
Operator : MD/SY
Sample : 05806-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E0LR3

Quant Time: Dec 15 00:52:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 15 00:49:12 2023
Response via : Initial Calibration









#13
Acetone
Concen: 5.232 ug/L
RT: 2.634 min Scan# 418
Delta R.T. -0.000 min
Lab File: VU056837.D
Acq: 14 Dec 2023 14:29




Instrument :
MSVOA_U
ClientSampleId :
E0LR3


Tgt Ion: 43 Resp: 12978
Ion Ratio Lower Upper
43 100
58 15.7 0.0 72.6



#42
Toluene
Concen: 0.268 ug/L
RT: 7.968 min Scan# 2077
Delta R.T. 0.003 min
Lab File: VU056837.D
Acq: 14 Dec 2023 14:29

Tgt Ion: 91 Resp: 24053
Ion Ratio Lower Upper
91 100
92 50.7 40.4 75.0





#53
 m,p-Xylene
 Concen: 0.258 ug/L
 RT: 9.695 min Scan# 20
 Delta R.T. 0.003 min
 Lab File: VU056837.D
 Acq: 14 Dec 2023 14:29

Instrument :
 MSVOA_U
 ClientSampleId :
 E0LR3

Tgt Ion:106 Resp: 9522
 Ion Ratio Lower Upper
 106 100
 91 187.2 154.0 286.0

