

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121624\
 Data File : VU062344.D
 Acq On : 16 Dec 2024 15:12
 Operator : MD/SY
 Sample : P5228-09
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2AT1

Quant Time: Dec 17 00:34:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 17 00:31:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.241	114	93381	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	92573	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	42559	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	32076	5.092	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 101.800%		
7) Chloroethane-d5	1.910	69	25341	5.656	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 113.200%		
11) 1,1-Dichloroethene-d2	2.560	65	15808	5.276	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 105.600%		
20) 2-Butanone-d5	4.631	46	58447	48.418	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 96.840%		
24) Chloroform-d	5.052	84	70729	5.189	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 103.800%		
26) 1,2-Dichloroethane-d4	5.695	65	38167	5.694	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 113.800%		
32) Benzene-d6	5.717	84	130526	4.896	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 98.000%		
36) 1,2-Dichloropropane-d6	6.679	67	37791	4.954	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 99.000%		
41) Toluene-d8	7.891	98	119052	4.727	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 94.600%		
43) trans-1,3-Dichloroprop...	8.174	79	16034	4.775	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 95.400%		
46) 2-Hexanone-d5	8.627	63	46551	45.951	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 91.900%		
56) 1,1,2,2-Tetrachloroeth...	10.746	84	29160	4.912	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 98.200%		
66) 1,2-Dichlorobenzene-d4	12.183	152	42681	5.442	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 108.800%		
Target Compounds						
16) Methylene chloride	3.039	84	6211	0.938	ug/L	94
25) Chloroform	5.074	83	6181	0.462	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121624\
Data File : VU062344.D
Acq On : 16 Dec 2024 15:12
Operator : MD/SY
Sample : P5228-09
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2AT1

Quant Time: Dec 17 00:34:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 17 00:31:28 2024
Response via : Initial Calibration

