

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121624\
 Data File : VU062370.D
 Acq On : 17 Dec 2024 05:38
 Operator : MD/SY
 Sample : P5262-01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2AS8

Quant Time: Dec 17 05:52:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 17 01:28:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	97064	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	91263	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	42588	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	30302	4.628	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	92.600%	
7) Chloroethane-d5	1.907	69	23813	5.114	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	102.200%	
11) 1,1-Dichloroethene-d2	2.560	65	14777	4.744	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	94.800%	
20) 2-Butanone-d5	4.631	46	57554	45.870	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	91.740%	
24) Chloroform-d	5.052	84	68219	4.814	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.200%	
26) 1,2-Dichloroethane-d4	5.692	65	36634	5.258	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.200%	
32) Benzene-d6	5.717	84	125437	4.772	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.400%	
36) 1,2-Dichloropropane-d6	6.682	67	35191	4.679	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.600%	
41) Toluene-d8	7.891	98	113910	4.587	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.800%	
43) trans-1,3-Dichloroprop...	8.174	79	14596	4.409	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.200%	
46) 2-Hexanone-d5	8.627	63	45049	45.107	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.220%	
56) 1,1,2,2-Tetrachloroeth...	10.746	84	28655	4.896	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.000%	
66) 1,2-Dichlorobenzene-d4	12.187	152	41509	5.289	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.800%	
Target Compounds						Qvalue
16) Methylene chloride	3.036	84	5675	0.825	ug/L	92
25) Chloroform	5.081	83	5090	0.366	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121624\
Data File : VU062370.D
Acq On : 17 Dec 2024 05:38
Operator : MD/SY
Sample : P5262-01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2AS8

Quant Time: Dec 17 05:52:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 17 01:28:58 2024
Response via : Initial Calibration

