

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121624\
 Data File : VU062366.D
 Acq On : 17 Dec 2024 04:00
 Operator : MD/SY
 Sample : VIBLK189
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK189

Quant Time: Dec 17 04:26:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 17 01:28:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	85081	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	78432	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	34618	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.593	65	26031	4.536	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	90.800%	
7) Chloroethane-d5	1.908	69	21753	5.329	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	106.600%	
11) 1,1-Dichloroethene-d2	2.560	65	12293	4.503	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	90.000%	
20) 2-Butanone-d5	4.631	46	57923	52.665	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	105.340%	
24) Chloroform-d	5.052	84	62401	5.024	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.400%	
26) 1,2-Dichloroethane-d4	5.692	65	35320	5.783	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.600%	
32) Benzene-d6	5.718	84	108872	4.820	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.400%	
36) 1,2-Dichloropropane-d6	6.679	67	33608	5.200	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.000%	
41) Toluene-d8	7.891	98	94208	4.415	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.200%	
43) trans-1,3-Dichloroprop...	8.174	79	13588	4.776	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	95.600%	
46) 2-Hexanone-d5	8.628	63	44458	51.797	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	103.600%	
56) 1,1,2,2-Tetrachloroeth...	10.747	84	26936	5.356	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.200%	
66) 1,2-Dichlorobenzene-d4	12.187	152	33960	5.323	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121624\
Data File : VU062366.D
Acq On : 17 Dec 2024 04:00
Operator : MD/SY
Sample : VIBLK189
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK189

Quant Time: Dec 17 04:26:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 17 01:28:58 2024
Response via : Initial Calibration

