

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121624\
 Data File : VU062368.D
 Acq On : 17 Dec 2024 04:49
 Operator : MD/SY
 Sample : P5253-23
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7J33

Quant Time: Dec 17 05:45:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 17 01:28:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	84583	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	79025	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	34693	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	26283	4.607	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.200%
7) Chloroethane-d5	1.907	69	21222	5.230	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.600%
11) 1,1-Dichloroethene-d2	2.560	65	11915	4.390	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.800%
20) 2-Butanone-d5	4.628	46	58412	53.423	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.840%
24) Chloroform-d	5.052	84	63294	5.126	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.600%
26) 1,2-Dichloroethane-d4	5.692	65	34983	5.762	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.200%
32) Benzene-d6	5.718	84	111039	4.879	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
36) 1,2-Dichloropropane-d6	6.682	67	33892	5.204	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.000%
41) Toluene-d8	7.891	98	95722	4.452	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.000%
43) trans-1,3-Dichloroprop...	8.174	79	13855	4.833	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.600%
46) 2-Hexanone-d5	8.627	63	44134	51.034	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.060%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26217	5.174	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	34155	5.342	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121624\
Data File : VU062368.D
Acq On : 17 Dec 2024 04:49
Operator : MD/SY
Sample : P5253-23
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7J33

Quant Time: Dec 17 05:45:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 17 01:28:58 2024
Response via : Initial Calibration

