

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121824\
 Data File : VU062425.D
 Acq On : 18 Dec 2024 16:32
 Operator : MD/SY
 Sample : VU1218WBL02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK119

Quant Time: Dec 19 00:14:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 19 00:12:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	95161	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	88430	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	40369	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	20793	3.239	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.800%
7) Chloroethane-d5	1.907	69	16955	3.714	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.200%
11) 1,1-Dichloroethene-d2	2.557	65	10736	3.516	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.400%
20) 2-Butanone-d5	4.627	46	47686	38.765	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.520%
24) Chloroform-d	5.052	84	56643	4.077	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.600%
26) 1,2-Dichloroethane-d4	5.695	65	30129	4.411	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
32) Benzene-d6	5.717	84	97368	3.823	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.400%
36) 1,2-Dichloropropane-d6	6.682	67	28885	3.964	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.200%
41) Toluene-d8	7.891	98	87846	3.651	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.000%
43) trans-1,3-Dichloroprop...	8.174	79	11892	3.707	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.200%
46) 2-Hexanone-d5	8.631	63	38818	40.113	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.220%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	23287	4.107	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	32578	4.379	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121824\
Data File : VU062425.D
Acq On : 18 Dec 2024 16:32
Operator : MD/SY
Sample : VU1218WBL02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK119

Quant Time: Dec 19 00:14:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 19 00:12:34 2024
Response via : Initial Calibration

