

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121824\
 Data File : VU062433.D
 Acq On : 18 Dec 2024 19:57
 Operator : MD/SY
 Sample : P5309-08DL 4X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CG3DL

Quant Time: Dec 19 00:16:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 19 00:12:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	73276	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	66672	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	29776	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	19852	4.016	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.400%
7) Chloroethane-d5	1.907	69	17084	4.860	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.200%
11) 1,1-Dichloroethene-d2	2.557	65	10005	4.255	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.200%
20) 2-Butanone-d5	4.634	46	50553	53.369	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.740%
24) Chloroform-d	5.052	84	56356	5.268	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.400%
26) 1,2-Dichloroethane-d4	5.695	65	30936	5.881	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.600%
32) Benzene-d6	5.718	84	95238	4.960	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
36) 1,2-Dichloropropane-d6	6.682	67	28857	5.252	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.000%
41) Toluene-d8	7.891	98	84869	4.679	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
43) trans-1,3-Dichloroprop...	8.174	79	12147	5.023	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.400%
46) 2-Hexanone-d5	8.631	63	38928	53.354	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.700%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	23884	5.586	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	33072	6.027	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	120.600%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121824\
Data File : VU062433.D
Acq On : 18 Dec 2024 19:57
Operator : MD/SY
Sample : P5309-08DL 4X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CG3DL

Quant Time: Dec 19 00:16:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 19 00:12:34 2024
Response via : Initial Calibration

