

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121824\
 Data File : VU062456.D
 Acq On : 19 Dec 2024 06:38
 Operator : MD/SY
 Sample : P5308-12
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CD3

Quant Time: Dec 19 08:09:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 19 01:28:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	74225	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	72869	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	34392	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18342	3.664	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.200%
7) Chloroethane-d5	1.907	69	16458	4.622	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.400%
11) 1,1-Dichloroethene-d2	2.557	65	9711	4.077	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.600%
20) 2-Butanone-d5	4.631	46	45037	46.938	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.880%
24) Chloroform-d	5.052	84	51837	4.784	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.600%
26) 1,2-Dichloroethane-d4	5.692	65	28087	5.271	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.400%
32) Benzene-d6	5.718	84	87744	4.181	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.600%
36) 1,2-Dichloropropane-d6	6.682	67	25939	4.320	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.400%
41) Toluene-d8	7.891	98	79827	4.026	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.600%
43) trans-1,3-Dichloroprop...	8.177	79	10358	3.919	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.400%
46) 2-Hexanone-d5	8.627	63	32584	40.861	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.720%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	21105	4.517	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	30614	4.831	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%
Target Compounds						
13) Acetone	2.653	43	2883	4.413	ug/L	98
34) Trichloroethene	6.537	95	4453	0.696	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121824\
Data File : VU062456.D
Acq On : 19 Dec 2024 06:38
Operator : MD/SY
Sample : P5308-12
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CD3

Quant Time: Dec 19 08:09:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 19 01:28:37 2024
Response via : Initial Calibration

