

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121924\
 Data File : VU062469.D
 Acq On : 19 Dec 2024 12:41
 Operator : MD/SY
 Sample : P5308-16DL 200X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CD7DL

Quant Time: Dec 20 00:37:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 20 00:36:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	69040	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	64402	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	27941	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	14840	3.187	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.800%
7) Chloroethane-d5	1.907	69	13969	4.217	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.400%
11) 1,1-Dichloroethene-d2	2.557	65	7761	3.503	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.000%
20) 2-Butanone-d5	4.627	46	43580	48.831	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.660%
24) Chloroform-d	5.052	84	45831	4.547	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
26) 1,2-Dichloroethane-d4	5.692	65	26236	5.294	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.800%
32) Benzene-d6	5.717	84	74723	4.029	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.600%
36) 1,2-Dichloropropane-d6	6.679	67	23658	4.458	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.200%
41) Toluene-d8	7.891	98	63250	3.610	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.200%
43) trans-1,3-Dichloroprop...	8.174	79	8972	3.841	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.800%
46) 2-Hexanone-d5	8.627	63	31436	44.605	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.200%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	20057	4.857	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	24041	4.669	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121924\
Data File : VU062469.D
Acq On : 19 Dec 2024 12:41
Operator : MD/SY
Sample : P5308-16DL 200X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CD7DL

Quant Time: Dec 20 00:37:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 20 00:36:00 2024
Response via : Initial Calibration

