

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121924\
 Data File : VU062479.D
 Acq On : 19 Dec 2024 16:46
 Operator : MD/SY
 Sample : P5309-19
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CL1

Quant Time: Dec 20 00:40:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 20 00:36:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	78420	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	73146	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	34869	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	18546	3.506	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.200%
7) Chloroethane-d5	1.908	69	15800	4.199	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.000%
11) 1,1-Dichloroethene-d2	2.560	65	9686	3.849	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.000%
20) 2-Butanone-d5	4.628	46	44568	43.965	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.920%
24) Chloroform-d	5.052	84	52083	4.550	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
26) 1,2-Dichloroethane-d4	5.692	65	29039	5.159	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
32) Benzene-d6	5.718	84	87828	4.169	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.400%
36) 1,2-Dichloropropane-d6	6.679	67	26379	4.376	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.600%
41) Toluene-d8	7.891	98	80967	4.068	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.400%
43) trans-1,3-Dichloroprop...	8.174	79	10683	4.026	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.600%
46) 2-Hexanone-d5	8.628	63	31951	39.916	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.840%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	20830	4.441	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	30747	4.785	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121924\
Data File : VU062479.D
Acq On : 19 Dec 2024 16:46
Operator : MD/SY
Sample : P5309-19
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CL1

Quant Time: Dec 20 00:40:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 20 00:36:00 2024
Response via : Initial Calibration

