

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121924\
 Data File : VU062495.D
 Acq On : 20 Dec 2024 02:18
 Operator : MD/SY
 Sample : P5308-14
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CD5

Quant Time: Dec 20 04:47:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 20 01:26:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	80894	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	77641	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	35023	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	21627	3.964	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.200%
7) Chloroethane-d5	1.907	69	17444	4.495	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.800%
11) 1,1-Dichloroethene-d2	2.560	65	10948	4.218	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.400%
20) 2-Butanone-d5	4.631	46	41045	39.251	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.500%
24) Chloroform-d	5.052	84	53277	4.512	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	5.692	65	28998	4.994	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.800%
32) Benzene-d6	5.717	84	92267	4.126	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.600%
36) 1,2-Dichloropropane-d6	6.679	67	26272	4.106	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.200%
41) Toluene-d8	7.891	98	86900	4.114	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.200%
43) trans-1,3-Dichloroprop...	8.174	79	11364	4.035	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.800%
46) 2-Hexanone-d5	8.627	63	29730	34.991	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	69.980%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	19972	4.011	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.200%
66) 1,2-Dichlorobenzene-d4	12.183	152	30871	4.783	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121924\
Data File : VU062495.D
Acq On : 20 Dec 2024 02:18
Operator : MD/SY
Sample : P5308-14
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CD5

Quant Time: Dec 20 04:47:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 20 01:26:48 2024
Response via : Initial Calibration

