

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122024\
 Data File : VU062533.D
 Acq On : 20 Dec 2024 18:42
 Operator : MD/SY
 Sample : P5345-13
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CM1

Quant Time: Dec 20 22:37:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 20 22:32:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	93573	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	89020	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	40689	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	23736	3.761	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.200%
7) Chloroethane-d5	1.907	69	18986	4.229	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.600%
11) 1,1-Dichloroethene-d2	2.560	65	12432	4.140	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.800%
20) 2-Butanone-d5	4.627	46	46684	38.594	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.180%
24) Chloroform-d	5.052	84	61156	4.477	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
26) 1,2-Dichloroethane-d4	5.692	65	32726	4.872	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.400%
32) Benzene-d6	5.717	84	103469	4.036	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.800%
36) 1,2-Dichloropropane-d6	6.682	67	29033	3.958	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.200%
41) Toluene-d8	7.888	98	97648	4.032	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.600%
43) trans-1,3-Dichloroprop...	8.177	79	13073	4.049	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.000%
46) 2-Hexanone-d5	8.627	63	36415	37.380	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.760%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	23029	4.034	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	36902	4.922	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.400%
Target Compounds					Qvalue	
9) Trichlorofluoromethane	2.132	101	2582	0.208	ug/L	92
13) Acetone	2.644	43	3681	4.469	ug/L	99
47) Tetrachloroethene	8.547	164	5247	0.870	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122024\
Data File : VU062533.D
Acq On : 20 Dec 2024 18:42
Operator : MD/SY
Sample : P5345-13
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CM1

Quant Time: Dec 20 22:37:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 20 22:32:26 2024
Response via : Initial Calibration

