

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122024\
 Data File : VU062523.D
 Acq On : 20 Dec 2024 14:35
 Operator : MD/SY
 Sample : P5308-17
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CD9

Quant Time: Dec 20 22:35:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 20 22:32:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	97146	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	90901	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	43314	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	23990	3.661	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.200%
7) Chloroethane-d5	1.907	69	20053	4.302	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.000%
11) 1,1-Dichloroethene-d2	2.557	65	12880	4.132	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.600%
20) 2-Butanone-d5	4.628	46	49479	39.401	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.800%
24) Chloroform-d	5.052	84	61736	4.353	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.000%
26) 1,2-Dichloroethane-d4	5.692	65	33385	4.787	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.800%
32) Benzene-d6	5.718	84	105229	4.019	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.400%
36) 1,2-Dichloropropane-d6	6.679	67	30152	4.025	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.600%
41) Toluene-d8	7.888	98	99911	4.040	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.800%
43) trans-1,3-Dichloroprop...	8.174	79	13465	4.084	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%
46) 2-Hexanone-d5	8.627	63	35975	36.165	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.320%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	23553	4.041	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	37419	4.688	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122024\
Data File : VU062523.D
Acq On : 20 Dec 2024 14:35
Operator : MD/SY
Sample : P5308-17
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CD9

Quant Time: Dec 20 22:35:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 20 22:32:26 2024
Response via : Initial Calibration

