

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122024\
 Data File : VU062526.D
 Acq On : 20 Dec 2024 15:49
 Operator : MD/SY
 Sample : P5345-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CK4

Quant Time: Dec 20 22:35:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 20 22:32:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	95065	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	90744	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	43893	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	21941	3.422	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.400%
7) Chloroethane-d5	1.907	69	18332	4.019	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.400%
11) 1,1-Dichloroethene-d2	2.557	65	11673	3.827	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.600%
20) 2-Butanone-d5	4.627	46	43669	35.535	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	71.080%
24) Chloroform-d	5.052	84	58232	4.196	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
26) 1,2-Dichloroethane-d4	5.692	65	31670	4.641	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.800%
32) Benzene-d6	5.717	84	100731	3.854	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.000%
36) 1,2-Dichloropropane-d6	6.682	67	28396	3.797	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	76.000%
41) Toluene-d8	7.891	98	94986	3.847	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.000%
43) trans-1,3-Dichloroprop...	8.174	79	13075	3.972	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.400%
46) 2-Hexanone-d5	8.627	63	32903	33.134	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	66.260%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	23032	3.958	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.200%
66) 1,2-Dichlorobenzene-d4	12.183	152	36201	4.476	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122024\
Data File : VU062526.D
Acq On : 20 Dec 2024 15:49
Operator : MD/SY
Sample : P5345-06
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CK4

Quant Time: Dec 20 22:35:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 20 22:32:26 2024
Response via : Initial Calibration

