

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU123024\
 Data File : VU062634.D
 Acq On : 30 Dec 2024 10:02
 Operator : MD/SY
 Sample : VU1230WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK132

Quant Time: Dec 31 00:17:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 31 00:16:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	99712	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	93648	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	45326	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	22202	3.301	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.000%
7) Chloroethane-d5	1.901	69	18863	3.943	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.800%
11) 1,1-Dichloroethene-d2	2.557	65	12189	3.809	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.200%
20) 2-Butanone-d5	4.618	46	70863	54.977	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.960%
24) Chloroform-d	5.049	84	65111	4.473	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
26) 1,2-Dichloroethane-d4	5.689	65	38831	5.425	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.600%
32) Benzene-d6	5.714	84	109443	4.058	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.200%
36) 1,2-Dichloropropane-d6	6.679	67	32585	4.222	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.400%
41) Toluene-d8	7.891	98	104691	4.109	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.200%
43) trans-1,3-Dichloroprop...	8.174	79	16651	4.902	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.000%
46) 2-Hexanone-d5	8.624	63	62860	61.338	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	122.680%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	29950	4.987	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	40726	4.876	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU123024\
Data File : VU062634.D
Acq On : 30 Dec 2024 10:02
Operator : MD/SY
Sample : VU1230WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK132

Quant Time: Dec 31 00:17:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 31 00:16:34 2024
Response via : Initial Calibration

