

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU123024\
 Data File : VU062635.D
 Acq On : 30 Dec 2024 10:26
 Operator : MD/SY
 Sample : VU1230WBL02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK133

Quant Time: Dec 31 00:17:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 31 00:16:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	99089	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	91146	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	44525	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	21843	3.268	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.400%
7) Chloroethane-d5	1.901	69	18675	3.928	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.600%
11) 1,1-Dichloroethene-d2	2.557	65	11153	3.508	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.200%
20) 2-Butanone-d5	4.615	46	59308	46.301	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.600%
24) Chloroform-d	5.052	84	62116	4.294	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.800%
26) 1,2-Dichloroethane-d4	5.692	65	36316	5.106	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%
32) Benzene-d6	5.714	84	106948	4.074	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.400%
36) 1,2-Dichloropropane-d6	6.679	67	30779	4.098	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.000%
41) Toluene-d8	7.891	98	102304	4.125	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
43) trans-1,3-Dichloroprop...	8.174	79	16047	4.854	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.000%
46) 2-Hexanone-d5	8.624	63	52499	52.634	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.260%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26927	4.607	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	39386	4.800	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU123024\
Data File : VU062635.D
Acq On : 30 Dec 2024 10:26
Operator : MD/SY
Sample : VU1230WBL02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK133

Quant Time: Dec 31 00:17:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 31 00:16:34 2024
Response via : Initial Calibration

